

No	Items	Technical Specification	Unit	Qty	Distribution to
1	OraQuick® HIV self-test	Tét nhanh phát hiện kháng thể đặc hiệu kháng HIV sử dụng dịch miệng. Độ nhạy >99% và độ đặc hiệu > 99.8%. Hạn sử dụng ít nhất 18 tháng kể từ khi nhận hàng. Đạt tiêu chuẩn tiền kiểm định của WHO. Mỗi test kit được đóng gói trong túi được gắn kín bao gồm nhưng không hạn chế các thành phần sau để phục vụ cho tự xét nghiệm: que thử và dụng cụ lấy mẫu, dung dịch đệm, giá để ống dung môi, tờ hướng dẫn sử dụng bằng tiếng Việt có hình ảnh minh họa (in màu). Rapid tests for qualitative detection of antibodies specific to HIV 1/2 in human oral liquid. Sensitivity >99% and specificity > 99.8%. Self-life is minimum 18 months from the date of receiving test kits. Pre-qualified by WHO. Each test kit is packaged in away that can be used for self testing Each test kit including but not limited to: testing device and specimen collection tool, buffer or diluent, stand, color printing of pictorial instruction in Vietnamese.	Test	2,000	Trung tâm Phòng chống HIV/AIDS Can Tho. Dương Nguyễn Văn Cu nói đại, phường An Bình, quận Ninh Kiều, TP Cần Thơ
	Total				